



CAR Guidelines and Standards for Cardiac Computed Tomography

The standards of the Canadian Association of Radiologists (CAR) are not rules, but are guidelines that attempt to define principles of practice that should generally produce radiological care. The physician and medical physicist may modify an existing standard as determined by the individual patient and available resources. Adherence to CAR standards will not assure a successful outcome in every situation. The standards should not be deemed inclusive of all proper methods of care or exclusive of other methods of care reasonably directed to obtaining the same results. The standards are not intended to establish a legal standard of care or conduct, and deviation from a standard does not, in and of itself, indicate or imply that such medical practice is below an acceptable level of care. The ultimate judgment regarding the propriety of any specific procedure or course of conduct must be made by the physician and medical physicist in light of all circumstances presented by the individual situation.

Approved: January 2009

These Standards were developed by the CAR Standards Working Group members: Carole J. Dennie, MD, Jonathon A. Leipsic, MD, Alan Brydie, MD.

Introduction

Invasive coronary angiography remains the gold standard for imaging of the coronary arteries. Because of poor temporal and spatial resolution, non-invasive imaging of the heart using computed tomography had remained a challenge until the recent past. Since 1999, and the advent of 4-detector ECG-gated CT, there have been rapid technical developments in CT technology and post-processing tools, thus enabling an accurate non-invasive assessment of cardiac anatomy including the coronary arteries as well as cardiac function. Today, this relatively new technique is being increasingly requested and performed on a routine basis.

Although guidelines and standards for the performance of cardiac CT (CCT) have been published by other societies outside of Canada [1-5], the Canadian Association of Radiologists recognizes that Canadian radiologists play a leading and pivotal role in the safe and proper implementation of CCT throughout the country, as well as in the training and continuing medical education of physicians performing and interpreting CCT studies. This comprehensive document reviews the current evidence for cardiac CT to date and outlines the standards for the implementation of a CCT program. Based on the review of the current literature and on expert opinion, recommendations as to indications and contraindications for CCT are also provided.

Methods

The cardiac CT expert committee comprises radiologists with cardiac expertise in each of the topic areas. Prior to completion, the standards and guidelines were distributed to the CAR executive for the opportunity to provide feedback concerning the recommendations. The literature will be periodically reviewed and the standards and guidelines will be updated as new or compelling evidence is identified.

Literature Search Strategy

The literature was searched using MEDLINE (OVID: 1966 through October 2008), EMBASE OVID: (1988 through October 2008), and the Cochrane Library (OVID: Issue 3, 2008). Reference lists of related papers and recent review articles were also scanned for additional citations.

Study Selection Criteria

Given the nature of the topic, it was widely accepted amongst the cardiac CT Writing Group that the strength of the evidence from the published literature would vary considerably, and in many cases would not be sufficient to inform recommendations on the topic. In the event of limited data, it was agreed that expert consensus would be used to form the recommendations. As such, only the highest levels of evidence were considered such as systematic reviews, randomized controlled trials, meta-analyses, non-randomized comparative studies, prospective single-cohort studies, and finally retrospective single-cohort studies. Articles were excluded from the systematic review of the evidence if they were reported in a language other than English or involved pediatric populations.

Results

Review of Cardiac CT Evidence

Calcium Score

Coronary artery calcium (CAC) is a general surrogate for total atheroma burden [6]. Most of the studies which have addressed CAC have been based on results from electron beam CT systems which are largely no longer available. These systems have been replaced by multidetector (MDCT). Early studies have demonstrated similar CAC scores can be obtained with these systems [7-9].

Outcome studies have shown that CAC scores add incremental prognostic value to the evaluation of asymptomatic patients at intermediate risk (10-20% 10 year risk) for a coronary event [10-15] in Caucasians using the Framingham risk score [16], the PROCAM score [17] or the European SCORE system [18]. The value of CAC scoring in asymptomatic patients at low risk or high risk of a coronary event is controversial [5, 19]. There is still limited data concerning the predictive value of CAC in non-Caucasians.

CAC is not an indicator of significant coronary artery stenosis. Investigators have concluded that although CAC scores are highly sensitive, they are only moderately specific for the detection of a coronary artery stenosis >50% [20].

A zero calcium score is associated with a very low event rate in most risk categories and is associated with a very low prevalence of ischemia on functional testing and significant coronary stenosis on invasive angiography [5, 21].

There is significant variability in CAC scores on MDCT on sequential examinations [22, 23].

Coronary Artery Imaging

Detection of Coronary Artery Stenosis

4 and 16 Detector CT

Non-invasive computed tomography imaging of the coronary arteries (CCTA) requires high temporal and spatial resolution and became possible with the development of multidetector (MDCT) technology.

The first generation of MDCT scanners were 4 slice systems. They were limited by long scan times, and lacked the temporal and spatial resolution of subsequent generations of scanners. The coronary arteries could be imaged but assessment was restricted to the proximal vessels and up to 25% of coronary segments were uninterpretable due to poor image quality [24, 25].

The subsequent 16-slice scanners had improved temporal and spatial resolution and acquired data in a shorter breathhold resulting in improved imaging of the coronary tree. A recent meta-analysis of papers comparing coronary CTA with conventional coronary angiography for detection of coronary stenoses >50% showed the following figures for 16 detector CT [26]. Per coronary segment analysis: sensitivity 77%, specificity 91%, positive predictive value 60% and negative predictive value 96%. Per patient analysis: sensitivity 95%, specificity 69%, positive predictive value 79% and negative predictive value 92%.

Despite the technical improvements with 16 detector CT, 4.4% of patients have non-evaluable scans and up to 29% (mean 10%) of coronary segments remain unassessable [26]. Exclusion of these unassessable patients and segments from analysis in many papers gives a false impression of the diagnostic performance of MDCT and the above figures must be interpreted with this in mind.

64 Detector CT

At the time of writing, the current generation of MDCT scanners are 64 detector technology. Having further improvements in spatial and temporal resolution and a shorter scan time, 64 detector CCTA allows significantly improved diagnostic performance over 16 detector technology [26-28]. Since the first publication on 64 detector CCTA in April 2005 [28], there has been a plethora of studies comparing 64 detector CCTA to conventional coronary angiography [29-41]. Five meta-analyses published over the last two years have had differing inclusion/exclusion criteria but confirm similar results, summarized in Tables 1 and 2 [26, 27, 42-44].

Per segment sensitivity is 88- 93%, specificity 96- 97%, positive predictive value 73-79% and negative predictive value 96-99%. Per patient sensitivity is 97-99%, specificity 88-93%, positive predictive value 93-94% and negative predictive value 95-100%. Non-evaluable scans occurred in 1.9% of patients and the mean unevaluable coronary segments was 4% [26].

The major advancements with 64 detector CCTA are a reduction in unevaluable scans and unevaluable vessel segments, and a considerable improvement in per patient specificity and positive predictive value. Negative predictive values are high for 16 and 64 detector for both per patient and per segment analyses.

Beyond Current 64 Detector Technology

At the time of writing, a new generation of CT scanners are being released. Equipment manufacturers are taking diverse paths with regards to advancing the technology. It is likely that these advancements will improve on the current 64 detector technology. There is considered insufficient literature concerning these new technologies to be included in this manuscript.

Disease Prevalence and Pre-test Cardiovascular Risk

The vast majority of studies included in the available meta-analyses are in patient groups with a high prevalence of coronary artery disease (mean prevalence of coronary disease 59% [26], 53% [42], 61% [44]). Disease prevalence has a bearing on the negative and positive predictive values of an investigation. High disease prevalence results in higher positive predictive value and lower negative predictive value, low prevalence results in lower positive predictive value and higher negative predictive value [45, 46]. Furthermore, although Bayesian theory dictates that the sensitivity or specificity of an investigation is not affected by disease prevalence, they are influenced by the composition of the population on which an investigation is assessed [47, 48]. In the context of CCTA, in a population with a low or intermediate likelihood of coronary disease, not only is the prevalence of coronary disease low, but it is to be expected that the severity of the disease will be less than in a high risk group, with different lesion composition and disease distribution within the coronary tree. The sensitivity and specificity of CCTA in the low and intermediate risk groups may therefore be different from that in the published meta-analyses.

Two studies have assessed the performance of CCTA in different patient groups. Hussman et al [48] stratified 88 patients into high, intermediate and low risk groups according to Framingham 10 year risk. Meijboom et al [49] stratified 254 patients into high, intermediate and low risk groups according to the Duke Clinical Score. Both studies found similar results. Specificity was lower in the high risk group. Positive predictive value was lower in low risk group. Negative predictive value was high across all groups in per patient and per segment analyses. The implication of these findings is that in low and intermediate risk groups, CCTA can reliably exclude disease but there will be an increasing number of false positive cases as pre-test risk decreases.

Table 1: 64 detector CCTA meta-analyses: per coronary segment analysis

	Sensitivity (%)	Specificity (%)	PPV (%)	NPV (%)
Vanhoenacker et al [27]	93	96	N/A	N/A
Hamon et al [26]	88	96	79	98
Sun et al [42]	90	96	75	98
Mowatt et al [43]	90	97	76	99
Stein et al [44]	90	96	73	96

Table 2: 64 detector CCTA meta-analyses: per patient analysis

	Sensitivity (%)	Specificity (%)	PPV (%)	NPV (%)
Vanhoenacker et al [27]	99	93	N/A	N/A
Hamon et al [26]	97	90	93	96
Sun et al [42]	97	88	94	95
Mowatt et al [43]	99	89	93	100
Stein et al [44]	98	88	93	96

Patient Parameters Influencing CCTA Diagnostic Performance

Certain patient parameters have been shown to adversely affect the diagnostic performance of 64-slice CCTA: coronary calcification, high heart rate, heart rate variability and body mass index.

Higher levels of coronary calcification are associated with poorer 64 detector CCTA diagnostic performance with various authors showing increased number of unassessable segments, lower specificity, lower positive predictive value and poorer image quality [32, 33, 50-52]. The widely accepted explanation for this is that calcification causes blooming artifact and beam hardening artifact which lead to an overestimation of the degree of stenosis. This effect is more pronounced when calcification is denser, however there is great heterogeneity amongst the studies in determining “high” and “low” calcification categories and there is no consensus as to an unacceptably high level for the performance of CCTA.

High heart rates are associated with poorer diagnostic performance of 64 detector CCTA [33] [51-55], due to motion artifact. In the majority of published papers, beta blocking medication was used to limit the heart rate. It should be noted that the target heart rate was variable (most commonly < 65 bpm) and not always achieved, but that in general, image quality improves as the heart rate is lower.

Variability of heart rate is associated with poorer diagnostic performance [55, 56] as well, due to data misregistration artifact between heart beats. Beta blockade is beneficial in reducing heart rate variability [55].

Obesity causes increased image noise which reduces contrast resolution of the coronary arteries [51]. Raff et al [33] reported decreased sensitivity, specificity, positive predictive value and negative predictive value in 64 detector CCTA in patients with a body mass index > 30 kg/m².

Multicentre Evidence

The vast body of evidence on the performance of 64 detector CCTA is based on single centre academic unit publications. Ong et al [41] reported findings from a centre with no prior experience in CCTA and showed very high negative predictive values but poorer per patient sensitivity and positive predictive values, and a higher percentage of unevaluable segments than those published by more experienced centres. They concluded that inexperienced centres may not be able to replicate the published experienced centre results.

At time of writing, the ACCURACY trial (30) is the only published multicentre trial assessing the performance of 64 detector CCTA. Two hundred and forty low to intermediate risk patients were recruited and had 64 detector CT and coronary angiography performed at 16 different centres, 83% of the patients being recruited and scanned at non-academic centres. CCTA studies were not read at the recruiting centre, but were read by two of three investigators, one of whom was from a non-academic centre. Per patient analysis for coronary artery stenoses greater or equal to 50% showed a sensitivity of 95%, specificity of 83%, positive predictive value of 64% and a negative predictive value of 99%. Similar to all the published single centre trials, negative predictive value was high, however sensitivity was slightly lower and positive predictive value was considerably lower. Whether this is due to the multicentre nature of the study or the relatively low prevalence of disease in the study population (25%) is unclear, however the positive predictive value is similar to that reported in studies assessing the performance of 64 detector CCTA in high versus low risk populations [48, 49].

Summary:

1. 64 detector CCTA outperforms 16 detector CCTA which in turn outperforms 4 detector CCTA. The major difference is in the number of unassessable studies and unassessable vessel segments.
2. 64 detector CCTA detects coronary stenoses of greater than or equal to 50% with a high sensitivity and high negative predictive value.
3. Positive predictive value is lower in populations with low disease prevalence.
4. Coronary calcification, high heart rate, variable heart rate and obesity have a negative impact on the diagnostic performance of 64 detector CCTA.
5. Results of inexperienced centres may not replicate those published by experienced academic centres.

Functional Relevance, Lesion Quantification and Characterization

It should be recognized that CCTA assesses the anatomy of the coronary tree and does not provide information as to the functional relevance of stenoses. Comparison of conventional coronary angiography with stress perfusion PET has shown that the vasodilator reserve (the ability to increase flow from baseline resting state) declines incrementally between 40% diameter stenosis up to 80% diameter stenosis [57]. Almost all of the studies of 64 detector CCTA have used the figure of 50% to represent a “significant” stenosis. When the functional relevance of these “significant stenoses” is assessed, a large proportion is found not to be associated with stress induced ischemia. Meijboom et al [58] found only 18% of CCTA stenoses \geq 50% to have a fractional flow reserve of < 0.75 , the level indicative of stress induced ischemia. In comparing 64-slice CCTA to SPECT myocardial perfusion imaging, Gaemperli et al [59] showed a CCTA stenosis \geq 50% to have only a 58% positive predictive value of positive myocardial perfusion imaging, Schuijff et al [60] showed only 39% of patients with stenosis \geq 50% to have abnormal myocardial perfusion imaging, and Scholte et al [61] showed 67% of patients with stenosis \geq 50% to have abnormal myocardial perfusion imaging.

A small number of studies have compared 64-slice CCTA to conventional angiography for detection of stenoses greater than 50%. Herzog et al [62] reported similar sensitivity, specificity, positive predictive value and negative predictive value for 50% and 70% lesions, with a negative predictive value of 100% for patient based analysis. Budoff et al (30) also reported very similar sensitivity, specificity and negative predictive value (99%) for 50% and 70% stenoses, however positive predictive value was lower for 70% lesions. Muhlenbruch et al [39] reported only on 70% lesions and showed a sensitivity of 98%, specificity of 50%, positive predictive value of 94% and negative predictive

value of 75%. In this study, prevalence of disease was 90%, which may explain the low negative predictive value and low specificity. Studies comparing 64-slice CCTA to conventional angiography that have categorized lesions further into quartile or smaller ranges of stenoses have shown a tendency for CTA to overestimate the degree of stenosis [63-65]. Raff et al [33] determined that although the mean difference between CCTA and conventional angiography grading of stenoses was small (1.3%), the standard deviation of differences was such that in only approximately 90% of cases, the CCTA grading was within +/- one quartile grading of the conventional angiogram.

Intravascular ultrasound (IVUS) is commonly used in the catheterization laboratory to give more anatomical information and more accurate quantification of stenotic coronary plaques [66]. Similar to IVUS, CCTA demonstrates not only the vessel lumen but also plaque and vessel wall. Several authors have compared 64 detector CCTA to IVUS for the detection and quantification of disease. Sun et al [67] reported excellent results for detection of plaque in a population with suspected coronary disease (sensitivity 97%, specificity 90%, positive predictive value 90%, negative predictive value 97%). In a population with less disease, Gregory et al [68] report poorer results for plaque detection (sensitivity 70%, specificity 92%, positive predictive value 89%, negative predictive value 77%). Several authors report reasonable correlation between CCTA and IVUS for lesion quantification. For percentage area stenosis Sato et al [69] report a correlation coefficient of 0.87 and Leber et al [30] report a correlation coefficient of 0.61. Both comment that CCTA tends to underestimate the percentage stenosis due to an overestimate of the lumen size. Caussin et al [70] report a correlation coefficient of 0.88 for assessment of mean luminal area. For the assessment of lesion plaque volume Leber et al [71] report a correlation coefficient of 0.69 and Otsuka et al [72] report a correlation coefficient of 0.98. In all of these studies, selected arteries or segments were used for the IVUS study and by the nature of the IVUS procedure, these assessments were generally made on larger proximal vessels.

Plaque characterization is an area of great interest. Several studies have reported 64-slice CCTA demonstration of certain characteristics which are more common in “culprit” lesions in patients with acute coronary syndrome such as positive remodelling, low attenuation plaque, spotty calcifications, ring-like foci [73-77]. Results for characterization of plaque composition are however disappointing in that although CCTA is excellent at differentiating calcified and non-calcified plaque, differentiation between the different components of non-calcified plaque (fibrous, fibrofatty, necrotic) is poor [67, 78].

Summary:

1. A significant percentage of coronary stenoses $\geq 50\%$ are not associated with ischemia.
2. There is little data regarding CCTA versus conventional angiographic quantification of lesions other than binary quantification around 50% stenosis. The data available indicates that there is considerable variability in quantification, however the tendency is for CCTA to overestimate lesions. Systematic overestimation of the degree of stenosis would maintain a high sensitivity and negative predictive value at the expense of specificity and positive predictive value.
3. On the basis of limited data, CCTA compares favourably with IVUS for detection of plaque in a population with high disease prevalence but less well when disease is less prevalent.
4. CCTA quantification of disease compares favourably with IVUS but tends to underestimate the degree of stenosis.
5. CCTA performs well in the differentiation of calcified and non-calcified plaques but poorly between different types of non-calcified plaques.

Assessment of Coronary Stents

Early evidence for 64 detector CCTA visualization of coronary stents was discouraging. Maintz et al [79] assessed the lumen visibility of 68 different coronary stents in an ex-vivo static model and reported only 10 of the stents to allow $>66\%$ lumen visibility. Stent diameters ranged from 2.5 mm to 4 mm (majority 3 mm). In vivo, Sheth et al [80] reported 56% of 54 stents scanned within 48 hours of deployment to be unassessable due to artifact.

Despite these shortcomings, a number of authors have subsequently shown 64 detector CCTA to have a reasonably high negative predictive value for detection of in-stent stenosis ranging from 89-100% [81-90] (Table 3). The literature shows a reasonably high but variable sensitivity (75- 100%) and specificity (74-100%) and a wide range of positive predictive values (39-100%). Also variable is the percentage of stents regarded as unassessable (range 0-42%). Pooled data from these studies show 11.8% of stents to be unassessable and to have been excluded from analysis. Factors influencing lumen visibility and stent assessability are stent diameter, stent material, stent strut size and density, overlapping stents, heart rate and body mass index [79, 85, 87, 91]. In particular, assessability is poor for stents ≤ 3 mm in diameter [80, 83, 87].

Summary:

1. A significant percentage of coronary stents prove to be unassessable by 64 detector CCTA.
2. Adverse features for stent assessment are small size (< 3mm), dense stents with large struts, overlapping stents, high heart rate, heart rate variability and high body mass index.
3. Excluding unassessable stents, 64 detector CCTA has a high negative predictive value for detection of in-stent stenosis $\geq 50\%$.

Table 3: 64 detector CCTA studies of coronary in-stent stenosis (per stent analysis)

	Stents	Excluded	Stenoses	Sensitivity	Specificity	PPV	NPV
Hecht [81]	132	0	17	94	74	39	99
Carrabba [82]	87	0	13	84	97	92	97
Carbone [83]	88	21	16	75	86	71	89
Das [84]	110	13	32	97	88	78	91
Schuijf [85]	76	11	6	100	100	100	100
Oncel [86]	39	0	19	89	95	94	90
Rixe [87]	102	43	12	88	98	86	98
Cademartiri [88]	192	14	20	95	93	63	99
Ehara [89]	125	15	24	91	93	77	98
Rist [92]	46	1	8	75	89	67	94

Assessment of Coronary Artery Bypass Grafts

Coronary artery bypass grafts are less mobile, contain less calcification, and in the case of vein grafts are larger than the coronary arteries, and so should be well suited for assessment with 64 detector CCTA. Multiple authors have addressed this and report similar results, despite differences in graft types (arterial or venous) and different study exclusion criteria with regards to high heart rate or arrhythmia [93-100]. The results are summarized in Table 4.

Table 4: 64 detector CCTA studies of coronary artery bypass graft stenosis (per graft analysis for combined stenosis $\geq 50\%$ and occlusion)

	Grafts(A/V)	Unassessable	Stenoses	Sensitivity	Specificity	PPV	NPV
Malagutti [93]	109 (45/64)	0	49	100	98	98	100
Pache [94]	96 (23/73)	3	45	98	89	90	98
Dijkers [95]	69 (52/17)	4	17	100	99	94	100
Ropers [96]	138 (37/101)	0	54	100	94	92	100
Meyer [97]	406 (147/259)	9	116	97	97	93	99
Jabara [98]	147 (47/100)	20	42	95	100	100	98
Onuma [99]	146 (74/72)	8	10	97	98	94	99
Feuchtner [100]	70 (46/24)	0	14	85	95	80	96

Similar to CCTA of the native coronary arteries, CCTA of coronary artery bypass grafts has extremely high sensitivity and negative predictive value, close to 100%. Specificity and positive predictive value are also very high, but slightly less impressive. In all studies differentiating occlusion from stenosis, performance was better for occlusion than for stenosis. The small numbers of errors (nearly all false positives) and unassessable grafts were due almost exclusively to metallic clips adjacent to the graft, most often at the distal anastomosis, and these were more common in arterial grafts. The size of the graft target vessel [100], heart rate and arrhythmia [97] are other factors which influence diagnostic accuracy. In two of the studies, grafts were depicted at CCTA that were not visualized at conventional angiography.

In clinical practice the status of the native non-grafted coronaries and the grafted run-off vessels is essential information for decisions on revascularisation. In this patient group, the native coronaries, and in particular the graft run-off vessels, have a higher likelihood of being small, diseased and calcified and so are relatively unfavourable for CCTA assessment. This is confirmed by those studies reporting assessment of the native coronary arteries and run-off vessels showing significantly poorer performance than the body of literature for assessment of coronary disease in non-grafted patients [93, 95, 96, 99].

Summary:

1. 64 detector CCTA has excellent negative predictive value and very good positive predictive value for detection of coronary artery bypass graft stenosis $\geq 50\%$.
2. A small percentage of grafts are unassessable by 64 detector CCTA.
3. Adverse features for graft assessment are adjacent metallic clips, arterial grafts, small target vessels, high heart rate and heart rate variability.
4. 64 detector CCTA determination of the status of the run-off vessels and native coronary arteries is relatively poor.

Imaging of Coronary Anomalies

Although coronary anomalies are relatively rare conditions, a small proportion have the potential to cause ischemia, myocardial infarction and sudden death [101]. In young athletes, coronary artery anomalies are the second most common cause of sudden death due to structural heart disease [102]. The identification of the origin and course of aberrant coronary arteries by conventional angiography can be difficult [103]. Because of the three-dimensional nature of the data set, CCTA is very well suited to detect and define the anatomic course of coronary artery anomalies and their relationship to other cardiac and non-cardiac structures. A number of case reports and several research papers [104-107] have demonstrated that the CCTA analysis of coronary anatomy in these patients is straightforward and very reliable with accuracy close to 100%.

Non-coronary Cardiac Imaging

Ventricular Function

Left Ventricular Function

Using advanced post-processing methods, left ventricular functional parameters such as end-diastolic and end-systolic volumes, stroke volume, ejection fraction, myocardial mass and regional wall motion abnormalities can be assessed and have shown good agreement with echocardiography, mono- and biplane ventriculography and gated SPECT as well as MRI [108-120]. A recent meta-analysis of MDCT left ventricular function analysis compared to MRI in 252 patients showed a weighted average difference of $-1.7 \pm 3.1\%$, a difference that is not relevant in clinical practice [121].

Right Ventricular Function

There is limited data on the accuracy of MDCT in assessing right ventricular function. Right ventricular (RV) quantification requires optimized contrast opacification of the RV. Small studies, mostly using 16 detector CT have shown good correlation of RV ejection fraction, RV end-diastolic volume, RV end-systolic volume and stroke volume with radionuclide ventriculography, MRI, cardiac catheterization and echocardiography in patients suspected of coronary artery disease [116, 122, 123], suspected right ventricular dysfunction [124, 125], suspected pulmonary emboli [126, 127] and congenital heart disease [128].

Valvular Function

The assessment of aortic stenosis has been the subject of several studies using 16, 40 and 64 detector CT. Aortic valve area using planimetry has been compared to transthoracic echocardiography (TTE) using the Doppler continuity equation, cine MR planimetry and transesophageal echo planimetry in patients with and without aortic stenosis. All studies showed good correlation between CT and the other modalities ($r=0.76-0.99$) [92, 129-134]. Some authors found a slight systematic overestimation of aortic valve area compared to TTE [132, 133]. Only two authors evaluated the sensitivity and specificity of 64-detector CT in the detection of aortic stenosis with sensitivities ranging between 82-100% and specificities between 77-93.7% [129, 133]. In his study of 52 patients with aortic stenosis evaluated with 64-detector CT, Habis also found good interobserver agreement (difference=0.002, variability=0.112cm²) [129].

Fewer data exist concerning the use of CT in the evaluation of aortic regurgitation. All studies showed good correlation of aortic regurgitant area with the severity of aortic valve regurgitation on TTE ($r=0.75-0.86$) [135-137]. Feuchtner found good interobserver agreement ($r=0.97$) for the determination of the aortic regurgitant area [135]. However, both Feuchtner and Jassal found a low negative predictive value especially in patients with mild aortic regurgitation, possibly due to artifact caused by the presence of aortic valve calcification.

Finally, with regards to mitral valvular disease, only one study could be found comparing 16 -detector CT with TEE and catheter ventriculography in 19 patients with mitral regurgitation. CT planimetry of the mitral valve regurgitant orifice correlated significantly with the other modalities [138].

Myocardial Perfusion and Viability

While the pharmacokinetics of CT contrast agents will allow first-pass perfusion imaging, and assessment of delayed enhancement for myocardial viability, only preliminary studies are available showing good agreement with MR [109, 139-146]. The radiation dose required to perform these studies remains a concern.

Left Atrium and Pulmonary Vein Assessment

It has been shown that the pulmonary veins are the source of triggers initiating atrial fibrillation in 90 to 96% of patients and that these foci can be effectively eliminated using catheter ablation [147]. Success rates in patients without underlying structural heart disease is over 80% [148]. Different ablation techniques include ostial segmental isolation of the pulmonary vein [149] and anatomically based circumferential ablation [150]. Pre procedural knowledge of the left atrial and pulmonary vein anatomy is crucial for the electrophysiologist and this can be provided with multidetector CT. Four detector and higher scanners can characterize posterior left atrial and pulmonary vein anatomy accurately without and with ECG gating [151]. 16 and 64 detector CT offer the advantage of decreased scan time, decreased cardiac motion and isotropic data sets which improve image quality even without gating. Important information concerning the number, size, distance from the ostium to the first branch and the presence of anatomic variants of pulmonary veins are important in order to help select the size of the ablation catheters used to perform the procedure. The dimensions of the left atrium, the presence of left atrial appendage thrombus and the anatomic course of the esophagus relative to the posterior left atrial wall and pulmonary veins can also be assessed [152, 153].

Image integration systems for catheter ablation procedures are now being used. With this technology, the 3 dimensional CT reconstructions are merged with the electroanatomic mapping data at the time of the procedure with an accuracy of 2 mm distance between corresponding points on the two images [154]. Some authors have found an increased success rate for catheter ablation using this technique [155, 156]. Kistler et al also found a decrease in fluoroscopy time [155].

Finally, MDCT has proven to be useful in the follow-up of patients after ablation therapy in order to assess for the development of complications especially to monitor the development of pulmonary vein stenosis [157, 158].

Coronary Vein Anatomy

In cardiac resynchronization therapy (CRT), left ventricular (LV) pacing is achieved by positioning the LV lead in one of the tributaries of the coronary sinus (CS). Although the success rate for transvenous LV lead placement is relatively high, in 5% to 12% of patients, the procedure does not succeed [159], and these numbers may be even higher in inexperienced centers. Failure of LV lead placement has been attributed to the inability to insert catheters in the CS and the lack of suitable side branches [159, 160]. Knowledge of the cardiac venous anatomy before these procedures may facilitate LV lead positioning. In 2005, Jongloed et al showed that non-invasive visualization of the coronary venous anatomy was feasible with 16 detector CT [161] and Van de Veire et al showed that visualization of the major tributaries of the coronary sinus was comparable between invasive venography and MDCT venography [162]. He also suggested that an additional 2 second delay should be applied after the contrast bolus reached the descending aorta before triggering the scan would optimize the scan for venous visualization.

Congenital Heart Disease

The population of adults with congenital heart disease is rapidly increasing, as a result of improved outcomes of surgical and catheter-based treatment strategies. The most obvious and clear indication for CCTA selection over MR in imaging these patients is the presence of a permanent implanted pacemaker or automated implantable cardiac defibrillator. Cardiac CT should act to augment the data collected using echocardiography particularly when there are limitations to the echocardiographic examination due to poor acoustic windows in the setting of prior cardiac surgery or chest wall deformity [163]. CT also supplements echo data in areas of echocardiographic weakness, particularly in the evaluation of the aortic arch, coronary arteries, branch pulmonary arteries, and collateral vessels.

The decision to use CCTA should be based upon the question to be answered at hand [164]. CT strengths in congenital heart disease include but are not limited to its fast acquisition time limiting the need for sedation. Its inherent high-spatial resolution offers evaluation of cardiac chamber size, conduits, baffles, aortic arch, great vessels, and pulmonary arteries and veins [165].

Pericardial Disease and Cardiac Masses

Echocardiography is the modality of choice in the initial investigation of pericardial disease and cardiac masses. It provides high-resolution, real-time images with recently improved tissue characterization using tissue harmonics and contrast echo [166]. However, because of restricted imaging windows and limited tissue characterization, CT and MR play an important role in the evaluation of pericardial thickening and cardiac masses. CT is superior to MR in the detection of calcification and to evaluate the extra-cardiac extent of disease such as involvement of the lungs. CT is also faster and less operator dependent than either echo or MR. MR has much better soft tissue contrast and enhancement characteristics of masses can be assessed without the use of ionizing radiation. In addition, the physiologic effects of pericardial abnormalities on the cardiac chambers are better characterized with MR as compared to CT because of its higher temporal resolution [167-169].

Extra-cardiac Findings

As compared to echocardiography, nuclear imaging and conventional coronary angiography, cardiac CT is unique in its ability to image not only the heart, but also the surrounding mediastinum, pulmonary vasculature, lungs, chest wall and upper abdomen. Extra-cardiac findings are quite commonly found on cardiac CT examinations. Multiple investigators have reported on the incidence and significance of these findings on cardiac CTA and electron-beam CT for calcium scoring. Although the definition of significant findings varied between studies, 4.2% to 22.7% of patients were reported to have findings which required additional investigations or immediate intervention [170-178]. Haller et al showed that only 35.5% of the total chest volume was displayed on dedicated coronary artery MDCT focused on the heart, whereas 70.3% of the chest was visible when coronary artery MDCT raw data were reconstructed with the maximal field of view [172].

Discussion

Based on the literature review, and consensus expert opinion, guidelines and standards for the performance of cardiac CT are provided below.

Guidelines for the Performance of Cardiac CT

Calcium Score

The Writing Group supports the use of calcium scoring in asymptomatic patients with an intermediate risk of cardiovascular events using a traditional scoring system, as this may influence the decision to intensify risk factor modification.

The Writing Group does not support calcium scoring:

1. In asymptomatic patients at low or high risk for cardiovascular events.
2. To monitor CAC progression over time.

Coronary Artery Imaging

Coronary CTA should only be performed in centres with adequate equipment by adequately trained staff. Studies must be supervised and interpreted by adequately trained physicians (see Standards Section). Studies should only be performed on patients in whom a diagnostic quality study is likely to be obtained and in whom the result of the study will influence patient management.

Coronary Artery Evaluation: Clinically Stable Patients

The Writing Group supports the use of coronary CTA in:

1. Symptomatic patients with low to intermediate pre-test probability of obstructive coronary artery disease who would otherwise be considered for conventional coronary angiography. This would typically be patients with chest pain and an equivocal or uninterpretable stress test.
2. Patients at low to intermediate risk of coronary artery disease with planned surgery for valvular or structural heart disease who would otherwise require pre-operative conventional coronary angiography.

The Writing Group does not support the use of coronary CTA in:

1. Symptomatic patients with high pre-test probability of obstructive coronary artery disease or previously documented coronary artery disease.
2. Asymptomatic patients. It should be emphasized that there currently is no evidence to support the use of coronary CTA as a screening examination for coronary artery disease.

Coronary Artery Evaluation: Clinically Unstable Patients

The use of coronary CTA in acute chest pain is controversial. The high negative predictive value of coronary CTA is a valuable tool, but the relatively poorer positive predictive value, particularly in populations with a low disease prevalence, is a potentially problematic source of false positive studies. Furthermore there is relatively poor correlation between CTA detected obstructive lesions and myocardial ischemia. Therefore positive studies require further assessment with either stress testing and/or conventional coronary angiography.

The Writing Group advocates the use of coronary CTA in patients with acute chest pain only in collaboration with experienced clinicians for patients with low to intermediate pre-test probability of coronary artery disease.

The Writing Group does not support the use of coronary CTA in patients with acute chest pain who have either a high pre-test probability of obstructive coronary artery disease or ECG or cardiac enzyme evidence of acute coronary syndrome.

Coronary Stent Evaluation

The Writing Group does not support the routine use of coronary CTA for evaluation of coronary artery stent patency.

The Writing Group supports the use of coronary CTA for the evaluation of coronary stent patency only in collaboration with experienced physicians in select cases with low to intermediate probability of stent stenosis who would otherwise have conventional coronary angiography. Stents should have a diameter > 3mm. It is advised that all unassessable and positive cases have conventional angiographic confirmation of stent status.

Coronary Artery Bypass Graft Evaluation

The Writing Group supports the use of coronary CTA for coronary artery bypass graft evaluation only in collaboration with experienced clinicians in patients in whom the clinical question is restricted to graft patency. The relatively poor performance for assessment of run-off vessels in these patients is likely to impede management decisions regarding revascularisation.

Coronary Artery Anomaly Evaluation

The Writing Group supports the use of coronary CTA for the evaluation of suspected clinically relevant coronary anomalies.

Non-coronary Cardiac Imaging

Ventricular Function

The Writing Group supports the use of cardiac CT in the assessment of ventricular function:

1. When a retrospectively gated examination is obtained for other accepted clinical indications or
2. If this information cannot be obtained through the use of other imaging modalities which do not require the use of ionizing radiation such as echocardiography or magnetic resonance imaging.

Valvular Function

The Writing Group supports the use of cardiac CT in the assessment of valvular function

1. When a retrospectively gated examination is obtained for other accepted clinical indications or
2. If this information cannot be obtained through the use of other imaging modalities which do not require the use of ionizing radiation such as echocardiography or magnetic resonance imaging.

Myocardial Perfusion and Viability

Because of the limited data to date, the Writing Group does not support the routine use of CT for assessing myocardial perfusion and viability unless this is part of a research study.

Left Atrium and Pulmonary Vein Assessment

The Writing Group supports the use of either gated or ungated cardiac CT for the assessment of the left atrium and pulmonary veins:

1. Prior to atrial fibrillation ablation.
2. To assess suspected post procedural complications.

Coronary Vein Anatomy

The Working Group supports the use of cardiac CT for the preprocedural assessment of the coronary veins in consultation with an electrophysiologist in order to answer a specific question which will affect the management of an individual patient.

Congenital Heart Disease

The Writing Group advocates a team approach of specialists with interest and knowledge in pediatric and adult congenital heart disease, including clinical and interventional cardiologists, radiologists and echocardiographers. As with any new technology and diagnostic technique several questions should be asked prior to selection of CT as the imaging test of choice over more well established techniques such as MRI and echo.

1. Does cardiac CT have the ability to answer the clinical question?
2. Will the results affect clinical management to justify the ionizing radiation exposure?
3. Can another test without ionizing radiation answer the clinical question without greater difficulty?
4. Is ECG gating necessary and if so, can dose reduction strategies such as prospective gating be used to lower exposure?

While there are no established criteria for patient selection, the Writing Group advocates a balanced and thoughtful approach to cardiac CT in congenital heart disease as outlined above.

Pericardial Disease and Cardiac Masses

The Writing Group supports the use of Cardiac CT in the investigation of pericardial disease or cardiac masses when:

1. The findings on echo or MRI are inconclusive or
2. There is a contraindication to MRI such as the presence of a pacemaker, claustrophobia or the inability to tolerate the examination.
3. CT is required to complete the staging of a probable cardiac malignancy.

Extra-cardiac Findings

Investigators have determined that only one third of the total chest volume is displayed on the coned down field of view images of a cardiac CT but the relatively high radiation dose acquisition contains information on the entire thorax in the range of z-axis covered.

Because of this, the Writing Group firmly believes that it would be unethical to exclude these structures from interpretation and therefore a second reconstruction of the images on mediastinal and lung windows at full field of view to allow the same breadth of visualization as a regular chest CT should routinely be performed in every case. These images should be reviewed by a radiologist in order to provide the opportunity for an alternative diagnosis which may account for the patient's symptoms or detect important but clinically occult problems such as early stage lung cancer.

Standards for the Performance of Cardiac CT

CT Facility Requirements

For diagnostic quality cardiac CT, a CT scanner should meet or exceed the following specifications:

1. For contrast enhanced coronary CT angiography (CCTA) a scanner must be capable of achieving in-plane resolution of $<0.5 \times 0.5$ mm axial, z- axis spatial resolution of <1 mm longitudinal, and temporal resolution of <0.25 seconds.
2. Tube heat capacity that allows for a single >20 -second acquisition.
3. All active CT facilities must have dose reduction strategies in place. This should include but not be limited to ECG dose modulation and specific protocols for smaller patients. Ideally labs should also have the ability to acquire data with prospective gating or "step-and shoot" sequential axial scanning.
4. Minimum section thickness no greater than 3 mm for calcium score CT and no greater than 1.5 mm for CT angiography [179].

To allow for adequate contrast-enhanced CCT, a power injector capable of delivering a programmed volume of a contrast agent at a steady flow rate of at least 4 cc per second for a delivery of >300 mg of iodine per millilitre is necessary. The precise optimal concentration of the contrast to be used is controversial. A dual-chambered power injector is a requirement for adequate coronary artery visualization and for other non-coronary CT applications [180].

Workstation capabilities must allow the interpreting physician to perform all of the necessary post-processing and data manipulation to ensure a thorough interpretation. These should include, but not be limited to multiplanar reformats (MPR), advanced vessel analysis, and volume rendering.

Patient Preparation

Because CCTA should not be performed in patients with an irregular heart beat such as atrial fibrillation, an ECG should be available or obtained prior to the scan. Contrast injection rates will vary between 4-8 cc/sec, and therefore patients should have an 18-gauge catheter or larger inserted, preferably in a cubital vein. As most CCTA examinations will be performed on 64-detector CT scanners, heart rate control is imperative in order to obtain diagnostic quality examinations and to reduce radiation dose. A heart rate of ≤ 65 beats per minute is desirable for all patients. Physicians should be familiar with the dosage and administration of oral and intravenous beta-blockers and calcium channel blockers as well as with contraindications to their use and their side-effects. Physicians should also be familiar with the treatment of adverse reactions to these medications.

All patients undergoing CCT should ideally receive oral nitrates immediately prior to image acquisition. Physicians should be aware of the contraindications to their use and the treatment of adverse reactions if they occur [2, 181].

Radiation Control

As in all imaging, care must always be taken to ensure the patient receives the lowest radiation dose possible. The policy of ALARA must be at the very heart of any coronary CT facility. The director or interpreting physician must be familiar with all of the recent dose reduction strategies [182, 183]. Prospective gating or so-called step and shoot axial scanning should be considered when it could adequately answer the clinical question. The decision to use such techniques should be made with an understanding of its limitations such as the lack of functional data and the inability to scan patients with significant variability in heart rate. Other steps to limit patient dose including the use of breast shields, individualized kVp selection and limits on z-axis coverage should be implemented [184]. Substantial dose savings can and should be realized by lowering the tube voltage from the routine 120 kV to 100 or 80kV [185]. This also results in increases in the level of vascular attenuation. Pooled experience would suggest that in patients with a BMI < 25 that 100 kV can be used routinely with more than satisfactory results. Similarly, low kV scanning should be used when performing cardiac CT in adolescent and pediatric patients for suspicion of coronary anomalies and other congenital cardiovascular conditions [186]. Dose modulation should be used routinely, except when planning on evaluating valve disease in systole which should be undertaken when other techniques have already been considered.

Reporting Standards

Structured and complete reporting are key elements for a functioning and valuable service in all areas of medical imaging. Never is this truer than in coronary CT angiography. While formal reporting standards are not being proposed, there are a number of recommendations that should be considered.

1. **Patient data**- demographics, indications, diagnosis, background data (Framingham risk assessment), relevant clinical history, and consent.
2. **Technical data**- medications administered for rate control and coronary dilatation, acquisition parameters, reconstruction techniques, vitals and complications. The technical component of the report should refer to the type of gating used for the study. It should also refer to the contrast material used, both the type and volume.
3. **Results**- The first statement in the results component of a report should refer to the quality of the study. This includes an overview statement and some explanation for limitations or artifacts that were encountered. Following the technical quality of the study, the report should consist of a formal evaluation of non-cardiac, cardiac but non-coronary, and coronary findings.

Non-coronary cardiac findings should include a review of the great vessels, myocardium, cardiac chambers, pericardium and valvular disease. Data regarding function and valve assessment would be limited to helical retrospective acquisitions.

Coronary assessment should include but not be limited to a review of dominance including a descriptor of course and branching pattern. There should also be an overview statement regarding the presence or absence of plaque burden and

the type of plaque. For the reporting of stenosis, use of the American Heart Association is recommended but is not felt to be a requirement [185]. Consistent and accurate vessel labelling and description is of utmost importance. Common terminology used in conventional angiography reports describing lesions as proximal, mid and distal according to the main coronary and branch anatomy should be used. Vessel size and distribution of various coronary segments is important to help guide clinical decision making. In addition, Agatston or mass calcium scores should be reported when formal calcium scanning is undertaken.

Stenosis evaluation and visual quantification should be performed in all coronary CT angiograms when plaque is present. While the strength of coronary CT angiography is in the exclusion of disease in patients with low pre-test probability, [49, 52] attempts should be made to quantify stenosis understanding the limitations of stenosis grading in CT due to limitations in spatial resolution. Quantitative CCTA is not required but a consistent grading scale should be used. A quartile system or a five point grading scale [187] is recommended to help guide the referring physician regarding the need for further non-invasive testing or invasive angiography. Stenosis descriptors should also include comments regarding vascular remodelling and plaque density when possible. All CT angiography facilities should pursue conventional angiographic follow up when available to assess internal accuracy.

Training and Continuing Medical Education Requirements

The training and continuing medical education requirements for the performance of cardiac CT are difficult and controversial issues. Because CCT is currently being performed by both radiologists and cardiologists, it is important that adequate training take place in both subspecialties. It is also important that both subspecialties work in a collaborative environment. For these reasons, the Canadian Association of Radiologists and the Canadian Cardiovascular Society are exploring the possibility of developing training standards for radiologists and cardiologists which would be acceptable to both societies as part of a larger collaborative document on cardiac CT. The training requirements will apply to a scope of cardiac CT (CCT) practice which includes the contrast-enhanced evaluation of cardiac chambers, coronary vessels and coronary bypass grafts and the non-enhanced evaluation of coronary calcium. These requirements will not include approval necessary for other vascular or thoracic imaging.

Conclusion

Cardiac CT allows a rapid non-invasive assessment of cardiac anatomy and function. Radiologists must play a crucial role in the education of physicians referring patients for cardiac CT and in the proper implementation of a cardiac CT program within their own institutions. The Canadian Association of Radiologists recognizes the importance of setting appropriate standards and guidelines for this rapidly evolving imaging technique.

Acknowledgments:

We gratefully acknowledge the editorial comments provided to us by the Canadian Cardiovascular Society during the development of this document. We have incorporated their valued suggestions into our final manuscript.

References

1. Budoff, M.J., et al., *ACCF/AHA clinical competence statement on cardiac imaging with computed tomography and magnetic resonance: a report of the American College of Cardiology Foundation/American Heart Association/American College of Physicians Task Force on Clinical Competence and Training*. *J Am Coll Cardiol*, 2005. **46**(2): p. 383-402.
2. Jacobs, J.E., et al., *ACR practice guideline for the performance and interpretation of cardiac computed tomography (CT)*. *J Am Coll Radiol*, 2006. **3**(9): p. 677-85.
3. Bluemke, D.A., et al., *Noninvasive coronary artery imaging: magnetic resonance angiography and multidetector computed tomography angiography: a scientific statement from the American heart association committee on cardiovascular imaging and intervention of the council on cardiovascular radiology and intervention, and the councils on clinical cardiology and cardiovascular disease in the young*. *Circulation*, 2008. **118**(5): p. 586-606.
4. Schroeder, S., et al., *Cardiac computed tomography: indications, applications, limitations, and training requirements: report of a Writing Group deployed by the Working Group Nuclear Cardiology and Cardiac CT of the European Society of Cardiology and the European Council of Nuclear Cardiology*. *Eur Heart J*, 2008. **29**(4): p. 531-56.
5. Greenland, P., et al., *ACCF/AHA 2007 clinical expert consensus document on coronary artery calcium scoring by computed tomography in global cardiovascular risk assessment and in evaluation of patients with chest pain: a report of the American College of Cardiology Foundation Clinical Expert Consensus Task Force (ACCF/AHA Writing Committee to Update the 2000 Expert Consensus Document on Electron Beam Computed Tomography)*. *Circulation*, 2007. **115**(3): p. 402-26.
6. Sangiorgi, G., et al., *Arterial calcification and not lumen stenosis is highly correlated with atherosclerotic plaque burden in humans: a histologic study of 723 coronary artery segments using nondecalcifying methodology*. *J Am Coll Cardiol*, 1998. **31**(1): p. 126-33.
7. Becker, C.R., et al., *Coronary artery calcium measurement: agreement of multirow detector and electron beam CT*. *AJR Am J Roentgenol*, 2001. **176**(5): p. 1295-8.
8. Daniell, A.L., et al., *Concordance of coronary artery calcium estimates between MDCT and electron beam tomography*. *AJR Am J Roentgenol*, 2005. **185**(6): p. 1542-5.
9. Detrano, R.C., et al., *Coronary calcium measurements: effect of CT scanner type and calcium measure on rescan reproducibility--MESA study*. *Radiology*, 2005. **236**(2): p. 477-84.
10. Vliegenthart, R., et al., *Coronary calcification improves cardiovascular risk prediction in the elderly*. *Circulation*, 2005. **112**(4): p. 572-7.
11. Arad, Y., et al., *Prediction of coronary events with electron beam computed tomography*. *J Am Coll Cardiol*, 2000. **36**(4): p. 1253-60.
12. Greenland, P., et al., *Coronary artery calcium score combined with Framingham score for risk prediction in asymptomatic individuals*. *JAMA*, 2004. **291**(2): p. 210-5.
13. Kondos, G.T., et al., *Electron-beam tomography coronary artery calcium and cardiac events: a 37-month follow-up of 5635 initially asymptomatic low- to intermediate-risk adults*. *Circulation*, 2003. **107**(20): p. 2571-6.
14. LaMonte, M.J., et al., *Coronary artery calcium score and coronary heart disease events in a large cohort of asymptomatic men and women*. *Am J Epidemiol*, 2005. **162**(5): p. 421-9.
15. Taylor, A.J., et al., *Coronary calcium independently predicts incident premature coronary heart disease over measured cardiovascular risk factors: mean three-year outcomes in the Prospective Army Coronary Calcium (PACC) project*. *J Am Coll Cardiol*, 2005. **46**(5): p. 807-14.

16. Wilson, P.W., et al., *Prediction of coronary heart disease using risk factor categories*. *Circulation*, 1998. **97**(18): p. 1837-47.
17. Assmann, G., P. Cullen, and H. Schulte, *Simple scoring scheme for calculating the risk of acute coronary events based on the 10-year follow-up of the prospective cardiovascular Munster (PROCAM) study*. *Circulation*, 2002. **105**(3): p. 310-5.
18. Conroy, R.M., et al., *Estimation of ten-year risk of fatal cardiovascular disease in Europe: the SCORE project*. *Eur Heart J*, 2003. **24**(11): p. 987-1003.
19. Naghavi, M., et al., *From vulnerable plaque to vulnerable patient--Part III: Executive summary of the Screening for Heart Attack Prevention and Education (SHAPE) Task Force report*. *Am J Cardiol*, 2006. **98**(2A): p. 2H-15H.
20. O'Rourke, R.A., et al., *American College of Cardiology/American Heart Association Expert Consensus document on electron-beam computed tomography for the diagnosis and prognosis of coronary artery disease*. *Circulation*, 2000. **102**(1): p. 126-40.
21. Budoff, M.J., et al., *Assessment of coronary artery disease by cardiac computed tomography: a scientific statement from the American Heart Association Committee on Cardiovascular Imaging and Intervention, Council on Cardiovascular Radiology and Intervention, and Committee on Cardiac Imaging, Council on Clinical Cardiology*. *Circulation*, 2006. **114**(16): p. 1761-91.
22. Horiguchi, J., et al., *Variability of repeated coronary artery calcium measurements by 1.25-mm- and 2.5-mm-thickness images on prospective electrocardiograph-triggered 64-slice CT*. *Eur Radiol*, 2008. **18**(2): p. 209-16.
23. Groen, J.M., et al., *Calcium scoring using 64-slice MDCT, dual source CT and EBT: a comparative phantom study*. *Int J Cardiovasc Imaging*, 2008. **24**(5): p. 547-56.
24. Achenbach, S., et al., *Detection of coronary artery stenoses by contrast-enhanced, retrospectively electrocardiographically-gated, multislice spiral computed tomography*. *Circulation*, 2001. **103**(21): p. 2535-8.
25. Kopp, A.F., et al., *Non-invasive coronary angiography with high resolution multidetector-row computed tomography. Results in 102 patients*. *Eur Heart J*, 2002. **23**(21): p. 1714-25.
26. Hamon, M., R. Morello, and J.W. Riddell, *Coronary arteries: diagnostic performance of 16- versus 64-section spiral CT compared with invasive coronary angiography--meta-analysis*. *Radiology*, 2007. **245**(3): p. 720-31.
27. Vanhoenacker, P.K., et al., *Diagnostic performance of multidetector CT angiography for assessment of coronary artery disease: meta-analysis*. *Radiology*, 2007. **244**(2): p. 419-28.
28. Leschka, S., et al., *Accuracy of MSCT coronary angiography with 64-slice technology: first experience*. *Eur Heart J*, 2005. **26**(15): p. 1482-7.
29. Ropers, D., et al., *Usefulness of multidetector row spiral computed tomography with 64- x 0.6-mm collimation and 330-ms rotation for the noninvasive detection of significant coronary artery stenoses*. *Am J Cardiol*, 2006. **97**(3): p. 343-8.
30. Leber, A.W., et al., *Quantification of obstructive and nonobstructive coronary lesions by 64-slice computed tomography: a comparative study with quantitative coronary angiography and intravascular ultrasound*. *J Am Coll Cardiol*, 2005. **46**(1): p. 147-54.
31. Mollet, N.R., et al., *High-resolution spiral computed tomography coronary angiography in patients referred for diagnostic conventional coronary angiography*. *Circulation*, 2005. **112**(15): p. 2318-23.
32. Ong, T.K., et al., *Accuracy of 64-row multidetector computed tomography in detecting coronary artery disease in 134 symptomatic patients: influence of calcification*. *Am Heart J*, 2006. **151**(6): p. 1323 e1-6.
33. Raff, G.L., et al., *Diagnostic accuracy of noninvasive coronary angiography using 64-slice spiral computed tomography*. *J Am Coll Cardiol*, 2005. **46**(3): p. 552-7.

34. Schuijff, J.D., et al., *Diagnostic accuracy of 64-slice multislice computed tomography in the noninvasive evaluation of significant coronary artery disease*. Am J Cardiol, 2006. **98**(2): p. 145-8.
35. Weustink, A.C., et al., *Reliable high-speed coronary computed tomography in symptomatic patients*. J Am Coll Cardiol, 2007. **50**(8): p. 786-94.
36. Ehara, M., et al., *Diagnostic accuracy of 64-slice computed tomography for detecting angiographically significant coronary artery stenosis in an unselected consecutive patient population: comparison with conventional invasive angiography*. Circ J, 2006. **70**(5): p. 564-71.
37. Nikolaou, K., et al., *Accuracy of 64-MDCT in the diagnosis of ischemic heart disease*. AJR Am J Roentgenol, 2006. **187**(1): p. 111-7.
38. Meijboom, W.B., et al., *Pre-operative computed tomography coronary angiography to detect significant coronary artery disease in patients referred for cardiac valve surgery*. J Am Coll Cardiol, 2006. **48**(8): p. 1658-65.
39. Muhlenbruch, G., et al., *Diagnostic value of 64-slice multi-detector row cardiac CTA in symptomatic patients*. Eur Radiol, 2007. **17**(3): p. 603-9.
40. Plass, A., et al., *Coronary artery imaging with 64-slice computed tomography from cardiac surgical perspective*. Eur J Cardiothorac Surg, 2006. **30**(1): p. 109-16.
41. Ong, K., et al., *Feasibility and accuracy of 64-row MDCT coronary imaging from a centre with early experience: a review and comparison with established centres*. Med J Malaysia, 2005. **60**(5): p. 629-36.
42. Sun, Z., et al., *Diagnostic value of 64-slice CT angiography in coronary artery disease: a systematic review*. Eur J Radiol, 2008. **67**(1): p. 78-84.
43. Mowatt, G., et al., *64-Slice computed tomography angiography in the diagnosis and assessment of coronary artery disease: systematic review and meta-analysis*. Heart, 2008. **94**(11): p. 1386-93.
44. Stein, P.D., et al., *64-slice CT for diagnosis of coronary artery disease: a systematic review*. Am J Med, 2008. **121**(8): p. 715-25.
45. Phillips, W.C., J.A. Scott, and G. Blasczynski, *Statistics for diagnostic procedures. I. How sensitive is "sensitivity"; how specific is "specificity"?* AJR Am J Roentgenol, 1983. **140**(6): p. 1265-70.
46. Altman, D.G. and J.M. Bland, *Diagnostic tests 2: Predictive values*. BMJ, 1994. **309**(6947): p. 102.
47. Brenner, H. and O. Gefeller, *Variation of sensitivity, specificity, likelihood ratios and predictive values with disease prevalence*. Stat Med, 1997. **16**(9): p. 981-91.
48. Husmann, L., et al., *Comparison of diagnostic accuracy of 64-slice computed tomography coronary angiography in patients with low, intermediate, and high cardiovascular risk*. Acad Radiol, 2008. **15**(4): p. 452-61.
49. Meijboom, W.B., et al., *64-slice computed tomography coronary angiography in patients with high, intermediate, or low pretest probability of significant coronary artery disease*. J Am Coll Cardiol, 2007. **50**(15): p. 1469-75.
50. Musto, C., et al., *64-multislice computed tomography in consecutive patients with suspected or proven coronary artery disease: initial single center experience*. Int J Cardiol, 2007. **114**(1): p. 90-7.
51. Brodoefel, H., et al., *Noninvasive coronary angiography using 64-slice spiral computed tomography in an unselected patient collective: effect of heart rate, heart rate variability and coronary calcifications on image quality and diagnostic accuracy*. Eur J Radiol, 2008. **66**(1): p. 134-41.
52. Budoff, M.J., et al., *Diagnostic Performance of 64-Multidetector Row Coronary Computed Tomographic Angiography for Evaluation of Coronary Artery Stenosis in Individuals Without Known Coronary Artery Disease: Results From the Prospective Multicenter ACCURACY (Assessment by Coronary Computed Tomographic Angiography of Individuals Undergoing Invasive Coronary Angiography) Trial*. Journal of the American College of Cardiology, 2008. **52**.

53. Delhaye, D., et al., *Coronary artery imaging during preoperative CT staging: preliminary experience with 64-slice multidetector CT in 99 consecutive patients*. Eur Radiol, 2007. **17**(3): p. 591-602.
54. Burgstahler, C., et al., *Quantitative parameters to compare image quality of non-invasive coronary angiography with 16-slice, 64-slice and dual-source computed tomography*. Eur Radiol, 2008.
55. Leschka, S., et al., *Noninvasive coronary angiography with 64-section CT: effect of average heart rate and heart rate variability on image quality*. Radiology, 2006. **241**(2): p. 378-85.
56. Leschka, S., et al., *Effect of decrease in heart rate variability on the diagnostic accuracy of 64-MDCT coronary angiography*. AJR Am J Roentgenol, 2008. **190**(6): p. 1583-90.
57. Uren, N.G., et al., *Relation between myocardial blood flow and the severity of coronary-artery stenosis*. N Engl J Med, 1994. **330**(25): p. 1782-8.
58. Meijboom, W.B., et al., *Comprehensive assessment of coronary artery stenoses: computed tomography coronary angiography versus conventional coronary angiography and correlation with fractional flow reserve in patients with stable angina*. J Am Coll Cardiol, 2008. **52**(8): p. 636-43.
59. Gaemperli, O., et al., *Functionally relevant coronary artery disease: comparison of 64-section CT angiography with myocardial perfusion SPECT*. Radiology, 2008. **248**(2): p. 414-23.
60. Schuijf, J.D., et al., *Relationship between noninvasive coronary angiography with multi-slice computed tomography and myocardial perfusion imaging*. J Am Coll Cardiol, 2006. **48**(12): p. 2508-14.
61. Scholte, A.J., et al., *Different manifestations of coronary artery disease by stress SPECT myocardial perfusion imaging, coronary calcium scoring, and multislice CT coronary angiography in asymptomatic patients with type 2 diabetes mellitus*. J Nucl Cardiol, 2008. **15**(4): p. 503-9.
62. Herzog, C., et al., *Significant coronary artery stenosis: comparison on per-patient and per-vessel or per-segment basis at 64-section CT angiography*. Radiology, 2007. **244**(1): p. 112-20.
63. Leber, A.W., et al., *Diagnostic accuracy of dual-source multi-slice CT-coronary angiography in patients with an intermediate pretest likelihood for coronary artery disease*. Eur Heart J, 2007. **28**(19): p. 2354-60.
64. Busch, S., et al., *Visual and automatic grading of coronary artery stenoses with 64-slice CT angiography in reference to invasive angiography*. Eur Radiol, 2007. **17**(6): p. 1445-51.
65. Husmann, L., et al., *Accuracy of quantitative coronary angiography with computed tomography and its dependency on plaque composition : Plaque composition and accuracy of cardiac CT*. Int J Cardiovasc Imaging, 2008. **24**(8): p. 895-904.
66. Tobis, J., B. Azarbal, and L. Slavin, *Assessment of intermediate severity coronary lesions in the catheterization laboratory*. J Am Coll Cardiol, 2007. **49**(8): p. 839-48.
67. Sun, J., et al., *Identification and quantification of coronary atherosclerotic plaques: a comparison of 64-MDCT and intravascular ultrasound*. AJR Am J Roentgenol, 2008. **190**(3): p. 748-54.
68. Gregory, S.A., et al., *Comparison of sixty-four-slice multidetector computed tomographic coronary angiography to coronary angiography with intravascular ultrasound for the detection of transplant vasculopathy*. Am J Cardiol, 2006. **98**(7): p. 877-84.
69. Sato, A., et al., *Quantitative measures of coronary stenosis severity by 64-Slice CT angiography and relation to physiologic significance of perfusion in nonobese patients: comparison with stress myocardial perfusion imaging*. J Nucl Med, 2008. **49**(4): p. 564-72.
70. Caussin, C., et al., *Comparison of coronary minimal lumen area quantification by sixty-four-slice computed tomography versus intravascular ultrasound for intermediate stenosis*. Am J Cardiol, 2006. **98**(7): p. 871-6.

71. Leber, A.W., et al., *Accuracy of 64-slice computed tomography to classify and quantify plaque volumes in the proximal coronary system: a comparative study using intravascular ultrasound*. J Am Coll Cardiol, 2006. **47**(3): p. 672-7.
72. Otsuka, M., et al., *Quantification of coronary plaque by 64-slice computed tomography: a comparison with quantitative intracoronary ultrasound*. Invest Radiol, 2008. **43**(5): p. 314-21.
73. Hoffmann, U., et al., *Noninvasive assessment of plaque morphology and composition in culprit and stable lesions in acute coronary syndrome and stable lesions in stable angina by multidetector computed tomography*. J Am Coll Cardiol, 2006. **47**(8): p. 1655-62.
74. Huang, W.C., et al., *Assessing culprit lesions and active complex lesions in patients with early acute myocardial infarction by multidetector computed tomography*. Circ J, 2008. **72**(11): p. 1806-13.
75. Nakazawa, G., et al., *Efficacy of culprit plaque assessment by 64-slice multidetector computed tomography to predict transient no-reflow phenomenon during percutaneous coronary intervention*. Am Heart J, 2008. **155**(6): p. 1150-7.
76. Motoyama, S., et al., *Multislice computed tomographic characteristics of coronary lesions in acute coronary syndromes*. J Am Coll Cardiol, 2007. **50**(4): p. 319-26.
77. Tanaka, A., et al., *Non-invasive assessment of plaque rupture by 64-slice multidetector computed tomography--comparison with intravascular ultrasound*. Circ J, 2008. **72**(8): p. 1276-81.
78. Choi, B.J., et al., *Comparison of 64-slice multidetector computed tomography with spectral analysis of intravascular ultrasound backscatter signals for characterizations of noncalcified coronary arterial plaques*. Am J Cardiol, 2008. **102**(8): p. 988-93.
79. Maintz, D., et al., *64-slice multidetector coronary CT angiography: in vitro evaluation of 68 different stents*. Eur Radiol, 2006. **16**(4): p. 818-26.
80. Sheth, T., et al., *Coronary stent assessability by 64 slice multi-detector computed tomography*. Catheter Cardiovasc Interv, 2007. **69**(7): p. 933-8.
81. Hecht, H.S., et al., *Usefulness of 64-detector computed tomographic angiography for diagnosing in-stent restenosis in native coronary arteries*. Am J Cardiol, 2008. **101**(6): p. 820-4.
82. Carrabba, N., et al., *Usefulness of 64-slice multidetector computed tomography for detecting drug eluting in-stent restenosis*. Am J Cardiol, 2007. **100**(12): p. 1754-8.
83. Carbone, I., et al., *Non-invasive evaluation of coronary artery stent patency with retrospectively ECG-gated 64-slice CT angiography*. Eur Radiol, 2008. **18**(2): p. 234-43.
84. Das, K.M., et al., *Contrast-enhanced 64-section coronary multidetector CT angiography versus conventional coronary angiography for stent assessment*. Radiology, 2007. **245**(2): p. 424-32.
85. Schuijff, J.D., et al., *Evaluation of patients with previous coronary stent implantation with 64-section CT*. Radiology, 2007. **245**(2): p. 416-23.
86. Oncel, D., G. Oncel, and M. Karaca, *Coronary stent patency and in-stent restenosis: determination with 64-section multidetector CT coronary angiography--initial experience*. Radiology, 2007. **242**(2): p. 403-9.
87. Rixe, J., et al., *Assessment of coronary artery stent restenosis by 64-slice multi-detector computed tomography*. Eur Heart J, 2006. **27**(21): p. 2567-72.
88. Cademartiri, F., et al., *Usefulness of 64-slice multislice computed tomography coronary angiography to assess in-stent restenosis*. J Am Coll Cardiol, 2007. **49**(22): p. 2204-10.
89. Ehara, M., et al., *Diagnostic accuracy of coronary in-stent restenosis using 64-slice computed tomography: comparison with invasive coronary angiography*. J Am Coll Cardiol, 2007. **49**(9): p. 951-9.

90. Rist, C., et al., *Assessment of coronary artery stent patency and restenosis using 64-slice computed tomography*. Acad Radiol, 2006. **13**(12): p. 1465-73.
91. Sun, Z., R. Davidson, and C.H. Lin, *Multi-detector row CT angiography in the assessment of coronary in-stent restenosis: A systematic review*. Eur J Radiol, 2007.
92. Pouleur, A.C., et al., *Aortic valve area assessment: multidetector CT compared with cine MR imaging and transthoracic and transesophageal echocardiography*. Radiology, 2007. **244**(3): p. 745-54.
93. Malagutti, P., et al., *Use of 64-slice CT in symptomatic patients after coronary bypass surgery: evaluation of grafts and coronary arteries*. Eur Heart J, 2007. **28**(15): p. 1879-85.
94. Pache, G., et al., *Initial experience with 64-slice cardiac CT: non-invasive visualization of coronary artery bypass grafts*. Eur Heart J, 2006. **27**(8): p. 976-80.
95. Dikkers, R., et al., *The benefit of 64-MDCT prior to invasive coronary angiography in symptomatic post-CABG patients*. Int J Cardiovasc Imaging, 2007. **23**(3): p. 369-77.
96. Ropers, D., et al., *Diagnostic accuracy of noninvasive coronary angiography in patients after bypass surgery using 64-slice spiral computed tomography with 330-ms gantry rotation*. Circulation, 2006. **114**(22): p. 2334-41; quiz 2334.
97. Meyer, T.S., et al., *Improved noninvasive assessment of coronary artery bypass grafts with 64-slice computed tomographic angiography in an unselected patient population*. J Am Coll Cardiol, 2007. **49**(9): p. 946-50.
98. Jabara, R., et al., *Comparison of multidetector 64-slice computed tomographic angiography to coronary angiography to assess the patency of coronary artery bypass grafts*. Am J Cardiol, 2007. **99**(11): p. 1529-34.
99. Onuma, Y., et al., *Evaluation of coronary artery bypass grafts and native coronary arteries using 64-slice multidetector computed tomography*. Am Heart J, 2007. **154**(3): p. 519-26.
100. Feuchtner, G.M., et al., *Diagnostic performance of 64-slice computed tomography in evaluation of coronary artery bypass grafts*. AJR Am J Roentgenol, 2007. **189**(3): p. 574-80.
101. Angelini, P., J.A. Velasco, and S. Flamm, *Coronary anomalies: incidence, pathophysiology, and clinical relevance*. Circulation, 2002. **105**(20): p. 2449-54.
102. Maron, B.J., *Sudden death in young athletes*. N Engl J Med, 2003. **349**(11): p. 1064-75.
103. Ishikawa, T. and P.W. Brandt, *Anomalous origin of the left main coronary artery from the right anterior aortic sinus: angiographic definition of anomalous course*. Am J Cardiol, 1985. **55**(6): p. 770-6.
104. Datta, J., et al., *Anomalous coronary arteries in adults: depiction at multi-detector row CT angiography*. Radiology, 2005. **235**(3): p. 812-8.
105. Duran, C., et al., *Remarkable anatomic anomalies of coronary arteries and their clinical importance: a multidetector computed tomography angiographic study*. J Comput Assist Tomogr, 2006. **30**(6): p. 939-48.
106. Kim, S.Y., et al., *Coronary artery anomalies: classification and ECG-gated multi-detector row CT findings with angiographic correlation*. Radiographics, 2006. **26**(2): p. 317-33; discussion 333-4.
107. Schmid, M., et al., *Visualization of coronary artery anomalies by contrast-enhanced multi-detector row spiral computed tomography*. Int J Cardiol, 2006. **111**(3): p. 430-5.
108. Schepis, T., et al., *Comparison of 64-slice CT with gated SPECT for evaluation of left ventricular function*. J Nucl Med, 2006. **47**(8): p. 1288-94.
109. Cury, R.C., et al., *Comprehensive assessment of myocardial perfusion defects, regional wall motion, and left ventricular function by using 64-section multidetector CT*. Radiology, 2008. **248**(2): p. 466-75.
110. Mahnken, A.H., et al., *Sixteen-slice spiral CT versus MR imaging for the assessment of left ventricular function in acute myocardial infarction*. Eur Radiol, 2005. **15**(4): p. 714-20.

111. Belge, B., et al., *Accurate estimation of global and regional cardiac function by retrospectively gated multidetector row computed tomography: comparison with cine magnetic resonance imaging*. *Eur Radiol*, 2006. **16**(7): p. 1424-33.
112. Dewey, M., et al., *Evaluation of global and regional left ventricular function with 16-slice computed tomography, biplane cineventriculography, and two-dimensional transthoracic echocardiography: comparison with magnetic resonance imaging*. *J Am Coll Cardiol*, 2006. **48**(10): p. 2034-44.
113. Henneman, M.M., et al., *Global and regional left ventricular function: a comparison between gated SPECT, 2D echocardiography and multi-slice computed tomography*. *Eur J Nucl Med Mol Imaging*, 2006. **33**(12): p. 1452-60.
114. Henneman, M.M., et al., *Assessment of global and regional left ventricular function and volumes with 64-slice MSCT: a comparison with 2D echocardiography*. *J Nucl Cardiol*, 2006. **13**(4): p. 480-7.
115. Muhlenbruch, G., et al., *Global left ventricular function in cardiac CT. Evaluation of an automated 3D region-growing segmentation algorithm*. *Eur Radiol*, 2006. **16**(5): p. 1117-23.
116. Puesken, M., et al., *Global left-ventricular function assessment using dual-source multidetector CT: effect of improved temporal resolution on ventricular volume measurement*. *Eur Radiol*, 2008. **18**(10): p. 2087-94.
117. Schlosser, T., et al., *Assessment of left ventricular function and mass in patients undergoing computed tomography (CT) coronary angiography using 64-detector-row CT: comparison to magnetic resonance imaging*. *Acta Radiol*, 2007. **48**(1): p. 30-5.
118. Schlosser, T., et al., *Assessment of left ventricular parameters using 16-MDCT and new software for endocardial and epicardial border delineation*. *AJR Am J Roentgenol*, 2005. **184**(3): p. 765-73.
119. Sugeng, L., et al., *Quantitative assessment of left ventricular size and function: side-by-side comparison of real-time three-dimensional echocardiography and computed tomography with magnetic resonance reference*. *Circulation*, 2006. **114**(7): p. 654-61.
120. van der Vleuten, P.A., et al., *Evaluation of global left ventricular function assessment by dual-source computed tomography compared with MRI*. *Eur Radiol*, 2008.
121. van der Vleuten, P.A., et al., *Quantification of global left ventricular function: comparison of multidetector computed tomography and magnetic resonance imaging. a meta-analysis and review of the current literature*. *Acta Radiol*, 2006. **47**(10): p. 1049-57.
122. Plumhans, C., et al., *Assessment of global right ventricular function on 64-MDCT compared with MRI*. *AJR Am J Roentgenol*, 2008. **190**(5): p. 1358-61.
123. Dogan, H., et al., *MDCT assessment of right ventricular systolic function*. *AJR Am J Roentgenol*, 2006. **186**(6 Suppl 2): p. S366-70.
124. Delhay, D., et al., *MDCT of right ventricular function: comparison of right ventricular ejection fraction estimation and equilibrium radionuclide ventriculography, part 1*. *AJR Am J Roentgenol*, 2006. **187**(6): p. 1597-604.
125. Remy-Jardin, M., et al., *MDCT of right ventricular function: impact of methodologic approach in estimation of right ventricular ejection fraction, part 2*. *AJR Am J Roentgenol*, 2006. **187**(6): p. 1605-9.
126. Coche, E., et al., *Evaluation of biventricular ejection fraction with ECG-gated 16-slice CT: preliminary findings in acute pulmonary embolism in comparison with radionuclide ventriculography*. *Eur Radiol*, 2005. **15**(7): p. 1432-40.
127. Dogan, H., et al., *Right ventricular function in patients with acute pulmonary embolism: analysis with electrocardiography-synchronized multi-detector row CT*. *Radiology*, 2007. **242**(1): p. 78-84.
128. Matsui, H., et al., *Quantification of right and left ventricular volumes in children with congenital heart disease by multidetector-row computed tomography*. *Pediatr Cardiol*, 2007. **28**(4): p. 267-71.
129. Feuchtner, G.M., et al., *Multislice computed tomography for detection of patients with aortic valve stenosis and quantification of severity*. *J Am Coll Cardiol*, 2006. **47**(7): p. 1410-7.

130. Feuchtner, G.M., et al., *Sixty-four slice CT evaluation of aortic stenosis using planimetry of the aortic valve area*. AJR Am J Roentgenol, 2007. **189**(1): p. 197-203.
131. Alkadhi, H., et al., *Aortic stenosis: comparative evaluation of 16-detector row CT and echocardiography*. Radiology, 2006. **240**(1): p. 47-55.
132. Tanaka, H., et al., *The simultaneous assessment of aortic valve area and coronary artery stenosis using 16-slice multidetector-row computed tomography in patients with aortic stenosis comparison with echocardiography*. Circ J, 2007. **71**(10): p. 1593-8.
133. Habis, M., et al., *Comparison of 64-slice computed tomography planimetry and Doppler echocardiography in the assessment of aortic valve stenosis*. J Heart Valve Dis, 2007. **16**(3): p. 216-24.
134. Bouvier, E., et al., *Diagnosis of aortic valvular stenosis by multislice cardiac computed tomography*. Eur Heart J, 2006. **27**(24): p. 3033-8.
135. Feuchtner, G.M., et al., *64-MDCT for diagnosis of aortic regurgitation in patients referred to CT coronary angiography*. AJR Am J Roentgenol, 2008. **191**(1): p. W1-7.
136. Jassal, D.S., et al., *64-slice multidetector computed tomography (MDCT) for detection of aortic regurgitation and quantification of severity*. Invest Radiol, 2007. **42**(7): p. 507-12.
137. Alkadhi, H., et al., *Aortic regurgitation: assessment with 64-section CT*. Radiology, 2007. **245**(1): p. 111-21.
138. Alkadhi, H., et al., *Mitral regurgitation: quantification with 16-detector row CT--initial experience*. Radiology, 2006. **238**(2): p. 454-63.
139. Nagao, M., et al., *Quantification of myocardial perfusion by contrast-enhanced 64-MDCT: characterization of ischemic myocardium*. AJR Am J Roentgenol, 2008. **191**(1): p. 19-25.
140. Nicol, E.D., et al., *Sixty-four-slice computed tomography coronary angiography compared with myocardial perfusion scintigraphy for the diagnosis of functionally significant coronary stenoses in patients with a low to intermediate likelihood of coronary artery disease*. J Nucl Cardiol, 2008. **15**(3): p. 311-8.
141. George, R.T., et al., *Quantification of myocardial perfusion using dynamic 64-detector computed tomography*. Invest Radiol, 2007. **42**(12): p. 815-22.
142. Kurata, A., et al., *Myocardial perfusion imaging using adenosine triphosphate stress multi-slice spiral computed tomography: alternative to stress myocardial perfusion scintigraphy*. Circ J, 2005. **69**(5): p. 550-7.
143. Nikolaou, K., et al., *Assessment of myocardial perfusion and viability from routine contrast-enhanced 16-detector-row computed tomography of the heart: preliminary results*. Eur Radiol, 2005. **15**(5): p. 864-71.
144. Lardo, A.C., et al., *Contrast-enhanced multidetector computed tomography viability imaging after myocardial infarction: characterization of myocyte death, microvascular obstruction, and chronic scar*. Circulation, 2006. **113**(3): p. 394-404.
145. Nieman, K., et al., *Reperfused myocardial infarction: contrast-enhanced 64-Section CT in comparison to MR imaging*. Radiology, 2008. **247**(1): p. 49-56.
146. Gerber, B.L., et al., *Characterization of acute and chronic myocardial infarcts by multidetector computed tomography: comparison with contrast-enhanced magnetic resonance*. Circulation, 2006. **113**(6): p. 823-33.
147. Shah, D.C., et al., *Electrophysiologically guided ablation of the pulmonary veins for the curative treatment of atrial fibrillation*. Ann Med, 2000. **32**(6): p. 408-16.
148. Oral, H., et al., *Pulmonary vein isolation for paroxysmal and persistent atrial fibrillation*. Circulation, 2002. **105**(9): p. 1077-81.
149. Morady, F., *Mechanisms and catheter ablation therapy of atrial fibrillation*. Tex Heart Inst J, 2005. **32**(2): p. 199-201.

150. Ouyang, F., et al., *Complete isolation of left atrium surrounding the pulmonary veins: new insights from the double-Lasso technique in paroxysmal atrial fibrillation*. *Circulation*, 2004. **110**(15): p. 2090-6.
151. Schwartzman, D., J. Lacomis, and W.G. Wigginton, *Characterization of left atrium and distal pulmonary vein morphology using multidimensional computed tomography*. *J Am Coll Cardiol*, 2003. **41**(8): p. 1349-57.
152. Cronin, P., et al., *MDCT of the left atrium and pulmonary veins in planning radiofrequency ablation for atrial fibrillation: a how-to guide*. *AJR Am J Roentgenol*, 2004. **183**(3): p. 767-78.
153. Lemola, K., et al., *Computed tomographic analysis of the anatomy of the left atrium and the esophagus: implications for left atrial catheter ablation*. *Circulation*, 2004. **110**(24): p. 3655-60.
154. Tops, L.F., et al., *Fusion of multislice computed tomography imaging with three-dimensional electroanatomic mapping to guide radiofrequency catheter ablation procedures*. *Heart Rhythm*, 2005. **2**(10): p. 1076-81.
155. Kistler, P.M., et al., *The impact of CT image integration into an electroanatomic mapping system on clinical outcomes of catheter ablation of atrial fibrillation*. *J Cardiovasc Electrophysiol*, 2006. **17**(10): p. 1093-101.
156. Martinek, M., et al., *Impact of integration of multislice computed tomography imaging into three-dimensional electroanatomic mapping on clinical outcomes, safety, and efficacy using radiofrequency ablation for atrial fibrillation*. *Pacing Clin Electrophysiol*, 2007. **30**(10): p. 1215-23.
157. Burgstahler, C., et al., *Visualization of pulmonary vein stenosis after radio frequency ablation using multi-slice computed tomography: initial clinical experience in 33 patients*. *Int J Cardiol*, 2005. **102**(2): p. 287-91.
158. Lacomis, J.M., et al., *CT of the pulmonary veins*. *J Thorac Imaging*, 2007. **22**(1): p. 63-76.
159. Abraham, W.T. and D.L. Hayes, *Cardiac resynchronization therapy for heart failure*. *Circulation*, 2003. **108**(21): p. 2596-603.
160. Puglisi, A., et al., *Limited thoracotomy as a second choice alternative to transvenous implant for cardiac resynchronisation therapy delivery*. *Eur Heart J*, 2004. **25**(12): p. 1063-9.
161. Jongbloed, M.R., et al., *Noninvasive visualization of the cardiac venous system using multislice computed tomography*. *J Am Coll Cardiol*, 2005. **45**(5): p. 749-53.
162. Van de Veire, N.R., et al., *Noninvasive imaging of cardiac venous anatomy with 64-slice multi-slice computed tomography and noninvasive assessment of left ventricular dyssynchrony by 3-dimensional tissue synchronization imaging in patients with heart failure scheduled for cardiac resynchronization therapy*. *Am J Cardiol*, 2008. **101**(7): p. 1023-9.
163. Cook, S.C., P.C. Dyke, and S.V. Raman, *Management of adults with congenital heart disease with cardiovascular computed tomography*. *Journal of Cardiovascular Computed Tomography*, 2008. **2**(1): p. 12-22.
164. Einstein, A.J., M.J. Henzlova, and S. Rajagopalan, *Estimating risk of cancer associated with radiation exposure from 64-slice computed tomography coronary angiography*. *JAMA*, 2007. **298**(3): p. 317-23.
165. Nicol, E.D., et al., *Assessment of adult congenital heart disease with multi-detector computed tomography: beyond coronary lumenography*. *Clin Radiol*, 2007. **62**(6): p. 518-27.
166. Thomas, J.D. and D.N. Rubin, *Tissue harmonic imaging: why does it work?* *J Am Soc Echocardiogr*, 1998. **11**(8): p. 803-8.
167. Breen, J.F., *Imaging of the pericardium*. *J Thorac Imaging*, 2001. **16**(1): p. 47-54.
168. Araoz, P.A., et al., *CT and MR imaging of benign primary cardiac neoplasms with echocardiographic correlation*. *Radiographics*, 2000. **20**(5): p. 1303-19.
169. Wang, Z.J., et al., *CT and MR imaging of pericardial disease*. *Radiographics*, 2003. **23 Spec No**: p. S167-80.
170. Kirsch, J., et al., *Prevalence and significance of incidental extracardiac findings at 64-multidetector coronary CTA*. *J Thorac Imaging*, 2007. **22**(4): p. 330-4.

171. Onuma, Y., et al., *Noncardiac findings in cardiac imaging with multidetector computed tomography*. J Am Coll Cardiol, 2006. **48**(2): p. 402-6.
172. Haller, S., et al., *Coronary artery imaging with contrast-enhanced MDCT: extracardiac findings*. AJR Am J Roentgenol, 2006. **187**(1): p. 105-10.
173. Schragin, J.G., et al., *Non-cardiac findings on coronary electron beam computed tomography scanning*. J Thorac Imaging, 2004. **19**(2): p. 82-6.
174. Hunold, P., et al., *Prevalence and clinical significance of accidental findings in electron-beam tomographic scans for coronary artery calcification*. Eur Heart J, 2001. **22**(18): p. 1748-58.
175. Horton, K.M., et al., *Prevalence of significant noncardiac findings on electron-beam computed tomography coronary artery calcium screening examinations*. Circulation, 2002. **106**(5): p. 532-4.
176. Mueller, J., et al., *Cardiac CT angiography after coronary bypass surgery: prevalence of incidental findings*. AJR Am J Roentgenol, 2007. **189**(2): p. 414-9.
177. Law, Y.M., et al., *Prevalence of significant extracoronary findings on multislice CT coronary angiography examinations and coronary artery calcium scoring examinations*. J Med Imaging Radiat Oncol, 2008. **52**(1): p. 49-56.
178. Dewey, M., et al., *Non-cardiac findings on coronary computed tomography and magnetic resonance imaging*. Eur Radiol, 2007. **17**(8): p. 2038-43.
179. Budoff, M.J., et al., *ACCF/AHA clinical competence statement on cardiac imaging with computed tomography and magnetic resonance*. Circulation, 2005. **112**(4): p. 598-617.
180. Kerl, J.M., et al., *Right heart: split-bolus injection of diluted contrast medium for visualization at coronary CT angiography*. Radiology, 2008. **247**(2): p. 356-64.
181. Kerl, J.M., et al., *Coronary CTA: image acquisition and interpretation*. J Thorac Imaging, 2007. **22**(1): p. 22-34.
182. Shuman, W.P., et al., *Prospective versus retrospective ECG gating for 64-detector CT of the coronary arteries: comparison of image quality and patient radiation dose*. Radiology, 2008. **248**(2): p. 431-7.
183. Earls, J.P., et al., *Prospectively gated transverse coronary CT angiography versus retrospectively gated helical technique: improved image quality and reduced radiation dose*. Radiology, 2008. **246**(3): p. 742-53.
184. Hausleiter, J., et al., *Radiation dose estimates from cardiac multislice computed tomography in daily practice: impact of different scanning protocols on effective dose estimates*. Circulation, 2006. **113**(10): p. 1305-10.
185. Schoepf, U.J., et al., *Coronary CT angiography*. Radiology, 2007. **244**(1): p. 48-63.
186. Siegel, M.J., et al., *Radiation dose and image quality in pediatric CT: effect of technical factors and phantom size and shape*. Radiology, 2004. **233**(2): p. 515-22.
187. Cheng, V., et al., *Moving beyond binary grading of coronary artery stenoses on coronary computed tomography: insights for the imager and referring clinician*. Journal of the American College of Cardiology Cardiovascular Imaging, 2008. **1**(4): p. 460-471.